



ENVIRONMENTAL DYNAMICS INC.
ATTN: Meighan Kearns
2195 - 2nd Avenue
Whitehorse YT Y1A 3T8

Date Received: 24-JUL-14
Report Date: 07-AUG-14 17:57 (MT)
Version: FINAL

Client Phone: 867-393-4882

Certificate of Analysis

Lab Work Order #: L1491715
Project P.O. #: NOT SUBMITTED
Job Reference: 14-Y-270
C of C Numbers: 1, 2
Legal Site Desc:

Can Dang
Senior Account Manager

[This report shall not be reproduced except in full without the written authority of the Laboratory.]

ADDRESS: 8081 Lougheed Hwy, Suite 100, Burnaby, BC V5A 1W9 Canada | Phone: +1 604 253 4188 | Fax: +1 604 253 6700
ALS CANADA LTD Part of the ALS Group A Campbell Brothers Limited Company

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID	L1491715-1 Water 21-JUL-14 16:00 NF2B	L1491715-2 Water 21-JUL-14 16:14 NF2	L1491715-3 Water 21-JUL-14 15:45 NF2A	L1491715-4 Water 21-JUL-14 15:21 X2-R	L1491715-5 Water 21-JUL-14 15:11 X2
Grouping	Analyte					
WATER						
Physical Tests	Conductivity (uS/cm)	165	180	173	181	183
	Hardness (as CaCO3) (mg/L)	85.3	88.2	90.4	92.3	91.2
	pH (pH)	7.94	7.85	7.87	7.93	7.95
	Total Suspended Solids (mg/L)	2.8	1.6	1.8	1.6	2.0
	Total Dissolved Solids (mg/L)	101	108	106	108	109
Anions and Nutrients	Alkalinity, Total (as CaCO3) (mg/L)	87.7	90.7	89.1	88.5	90.6
	Ammonia, Total (as N) (mg/L)	<0.0050	<0.0050	<0.0050	<0.0050	<0.0050
	Chloride (Cl) (mg/L)	<0.50	<0.50	<0.50	<0.50	<0.50
	Fluoride (F) (mg/L)	0.096	0.101	0.101	0.101	0.099
	Nitrate (as N) (mg/L)	0.0350	0.0361	0.0372	0.0328	0.0322
	Nitrite (as N) (mg/L)	<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Phosphorus (P)-Total (mg/L)	0.0058	0.0043	0.0036	0.0045	0.0037
	Sulfate (SO4) (mg/L)	11.8	15.7	14.7	16.4	16.4
	Anion Sum (meq/L)	2.00	2.15	2.09	2.12	2.16
	Cation Sum (meq/L)	1.81	1.91	1.93	1.97	1.94
	Cation - Anion Balance (%)	-5.1	-5.8	-4.1	-3.7	-5.2
Organic / Inorganic Carbon	Dissolved Organic Carbon (mg/L)	2.47	2.47	2.52	2.48	2.56
	Total Organic Carbon (mg/L)	2.72	2.63	2.61	2.53	2.58
Total Metals	Aluminum (Al)-Total (mg/L)	0.0678	0.0522	0.0530	0.0415	0.0424
	Antimony (Sb)-Total (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Arsenic (As)-Total (mg/L)	0.00083	0.00073	0.00074	0.00070	0.00069
	Barium (Ba)-Total (mg/L)	0.0512	0.0485	0.0541	0.0491	0.0486
	Beryllium (Be)-Total (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Bismuth (Bi)-Total (mg/L)	<0.00050	<0.00050	<0.00050	<0.00050	<0.00050
	Boron (B)-Total (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Cadmium (Cd)-Total (mg/L)	0.000016	0.000397	0.000175	0.000182	0.000183
	Calcium (Ca)-Total (mg/L)	25.1	25.8	25.5	26.5	26.9
	Chromium (Cr)-Total (mg/L)	0.00022	0.00037	0.00019	0.00017	0.00017
	Cobalt (Co)-Total (mg/L)	0.00012	0.00230	0.00098	0.00102	0.00103
	Copper (Cu)-Total (mg/L)	0.00076	0.00065	0.00069	0.00065	0.00064
	Iron (Fe)-Total (mg/L)	0.295	0.342	0.310	0.278	0.277
	Lead (Pb)-Total (mg/L)	0.00128	0.00115	0.00104	0.000917	0.000937
	Lithium (Li)-Total (mg/L)	0.00372	0.00353	0.00336	0.00401	0.00383
	Magnesium (Mg)-Total (mg/L)	5.27	6.37	5.79	6.06	6.07
	Manganese (Mn)-Total (mg/L)	0.0297	0.144	0.0710	0.0832	0.0839
	Molybdenum (Mo)-Total (mg/L)	0.000534	0.000567	0.000510	0.000538	0.000501

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID	L1491715-6 Water 21-JUL-14 14:32 X10	L1491715-7 Water 21-JUL-14 14:51 X3A	L1491715-8 Water 21-JUL-14 17:18 R10	L1491715-9 Water 21-JUL-14 16:57 NF1	L1491715-10 Water 21-JUL-14 17:45 FIELD BLANK
Grouping	Analyte					
WATER						
Physical Tests	Conductivity (uS/cm)	199	204	171	171	<2.0
	Hardness (as CaCO3) (mg/L)	106	102	87.0	88.8	<0.50
	pH (pH)	8.02	8.01	7.93	7.83	5.69
	Total Suspended Solids (mg/L)	<1.0	1.2	<1.0	2.6	<1.0
	Total Dissolved Solids (mg/L)	124	118	101	104	<1.0
Anions and Nutrients	Alkalinity, Total (as CaCO3) (mg/L)	101	97.3	88.3	89.9	<2.0
	Ammonia, Total (as N) (mg/L)	<0.0050	<0.0050	<0.0050	0.0076	<0.0050
	Chloride (Cl) (mg/L)	<0.50	<0.50	<0.50	<0.50	<0.50
	Fluoride (F) (mg/L)	0.106	0.104	0.097	0.096	<0.020
	Nitrate (as N) (mg/L)	0.0202	0.0250	0.0314	0.0312	<0.0050
	Nitrite (as N) (mg/L)	<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Phosphorus (P)-Total (mg/L)	0.0030	0.0026	0.0046	0.0024	<0.0020
	Sulfate (SO4) (mg/L)	18.9	17.4	11.4	12.3	<0.50
	Anion Sum (meq/L)	2.42	2.31	2.01	2.06	<0.10
	Cation Sum (meq/L)	2.25	2.15	1.84	1.89	<0.10
	Cation - Anion Balance (%)	-3.7	-3.7	-4.4	-4.2 ^{SP}	0.0
Organic / Inorganic Carbon	Dissolved Organic Carbon (mg/L)	2.57	2.56	2.57	2.46	<0.50
	Total Organic Carbon (mg/L)	2.77	2.55	2.62	2.88	<0.50
Total Metals	Aluminum (Al)-Total (mg/L)	0.0193	0.0243	0.0252	0.0841	<0.0030
	Antimony (Sb)-Total (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Arsenic (As)-Total (mg/L)	0.00050	0.00056	0.00072	0.00085	<0.00010
	Barium (Ba)-Total (mg/L)	0.0518	0.0509	0.0484	0.0513	<0.000050
	Beryllium (Be)-Total (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Bismuth (Bi)-Total (mg/L)	<0.00050	<0.00050	<0.00050	<0.00050	<0.00050
	Boron (B)-Total (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Cadmium (Cd)-Total (mg/L)	0.000082	0.000096	0.000014	0.000024	<0.000010
	Calcium (Ca)-Total (mg/L)	31.3	29.2	26.4	26.2	<0.020
	Chromium (Cr)-Total (mg/L)	<0.00010	0.00012	0.00013	0.00027	<0.00010
	Cobalt (Co)-Total (mg/L)	0.00040	0.00050	<0.00010	0.00026	<0.00010
	Copper (Cu)-Total (mg/L)	0.00064	0.00062	0.00059	0.00077	<0.00050
	Iron (Fe)-Total (mg/L)	0.417	0.237	0.185	0.431	<0.010
	Lead (Pb)-Total (mg/L)	0.000337	0.000597	0.000328	0.00210	<0.000050
	Lithium (Li)-Total (mg/L)	0.00305	0.00298	0.00354	0.00360	<0.00050
	Magnesium (Mg)-Total (mg/L)	7.31	6.71	5.44	5.46	<0.0050
	Manganese (Mn)-Total (mg/L)	0.0555	0.0590	0.0195	0.0767	<0.000050
	Molybdenum (Mo)-Total (mg/L)	0.000498	0.000433	0.000578	0.000540	<0.000050

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID				
	L1491715-11 Water TRAVEL BLANK				
Grouping	Analyte				
WATER					
Physical Tests	Conductivity (uS/cm)	<2.0			
	Hardness (as CaCO3) (mg/L)	<0.50			
	pH (pH)	5.65			
	Total Suspended Solids (mg/L)	<1.0			
	Total Dissolved Solids (mg/L)	<1.0			
Anions and Nutrients	Alkalinity, Total (as CaCO3) (mg/L)	<2.0			
	Ammonia, Total (as N) (mg/L)	<0.0050			
	Chloride (Cl) (mg/L)	<0.50			
	Fluoride (F) (mg/L)	<0.020			
	Nitrate (as N) (mg/L)	<0.0050			
	Nitrite (as N) (mg/L)	<0.0010			
	Phosphorus (P)-Total (mg/L)	<0.0020			
	Sulfate (SO4) (mg/L)	<0.50			
	Anion Sum (meq/L)	<0.10			
	Cation Sum (meq/L)	<0.10			
	Cation - Anion Balance (%)	0.0			
Organic / Inorganic Carbon	Dissolved Organic Carbon (mg/L)				
	Total Organic Carbon (mg/L)	<0.50			
Total Metals	Aluminum (Al)-Total (mg/L)	<0.0030			
	Antimony (Sb)-Total (mg/L)	<0.00010			
	Arsenic (As)-Total (mg/L)	<0.00010			
	Barium (Ba)-Total (mg/L)	<0.000050			
	Beryllium (Be)-Total (mg/L)	<0.00010			
	Bismuth (Bi)-Total (mg/L)	<0.00050			
	Boron (B)-Total (mg/L)	<0.010			
	Cadmium (Cd)-Total (mg/L)	<0.000010			
	Calcium (Ca)-Total (mg/L)	<0.020			
	Chromium (Cr)-Total (mg/L)	<0.00010			
	Cobalt (Co)-Total (mg/L)	<0.00010			
	Copper (Cu)-Total (mg/L)	<0.00050			
	Iron (Fe)-Total (mg/L)	<0.010			
	Lead (Pb)-Total (mg/L)	<0.000050			
	Lithium (Li)-Total (mg/L)	<0.00050			
	Magnesium (Mg)-Total (mg/L)	<0.0050			
	Manganese (Mn)-Total (mg/L)	<0.000050			
	Molybdenum (Mo)-Total (mg/L)	<0.000050			

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID	L1491715-1 Water 21-JUL-14 16:00 NF2B	L1491715-2 Water 21-JUL-14 16:14 NF2	L1491715-3 Water 21-JUL-14 15:45 NF2A	L1491715-4 Water 21-JUL-14 15:21 X2-R	L1491715-5 Water 21-JUL-14 15:11 X2
Grouping	Analyte					
WATER						
Total Metals	Nickel (Ni)-Total (mg/L)	0.00063	0.00376	0.00192	0.00192	0.00194
	Phosphorus (P)-Total (mg/L)	<0.30	<0.30	<0.30	<0.30	<0.30
	Potassium (K)-Total (mg/L)	0.582	0.580	0.600	0.594	0.592
	Selenium (Se)-Total (mg/L)	0.00028	0.00022	0.00024	0.00027	0.00023
	Silicon (Si)-Total (mg/L)	4.68	5.08	4.79	4.73	4.68
	Silver (Ag)-Total (mg/L)	0.000038	<0.000010	0.000020	0.000014	0.000013
	Sodium (Na)-Total (mg/L)	1.91	2.00	1.99	1.99	2.01
	Strontium (Sr)-Total (mg/L)	0.114	0.113	0.110	0.114	0.117
	Thallium (Tl)-Total (mg/L)	<0.000010	<0.000010	<0.000010	<0.000010	<0.000010
	Tin (Sn)-Total (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Titanium (Ti)-Total (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Uranium (U)-Total (mg/L)	0.00108	0.00108	0.00107	0.00111	0.00112
	Vanadium (V)-Total (mg/L)	<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Zinc (Zn)-Total (mg/L)	0.0085	0.595	0.241	0.253	0.255
	Zirconium (Zr)-Total (mg/L)	<0.00080	<0.00080	<0.00080	<0.00080	<0.00080
Dissolved Metals	Dissolved Metals Filtration Location	FIELD	FIELD	FIELD	FIELD	FIELD
	Aluminum (Al)-Dissolved (mg/L)	0.0068	0.0132	0.0086	0.0095	0.0097
	Antimony (Sb)-Dissolved (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Arsenic (As)-Dissolved (mg/L)	0.00059	0.00057	0.00063	0.00055	0.00055
	Barium (Ba)-Dissolved (mg/L)	0.0487	0.0489	0.0486	0.0483	0.0491
	Beryllium (Be)-Dissolved (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Bismuth (Bi)-Dissolved (mg/L)	<0.00050	<0.00050	<0.00050	<0.00050	<0.00050
	Boron (B)-Dissolved (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Cadmium (Cd)-Dissolved (mg/L)	0.000013	0.000566 ^{DTC}	0.000159	0.000168	0.000168
	Calcium (Ca)-Dissolved (mg/L)	25.5	24.2	26.1	27.0	26.6
	Chromium (Cr)-Dissolved (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Cobalt (Co)-Dissolved (mg/L)	<0.00010	0.00320 ^{DTC}	0.00089	0.00096	0.00097
	Copper (Cu)-Dissolved (mg/L)	0.00054	0.00051	0.00050	0.00050	0.00048
	Iron (Fe)-Dissolved (mg/L)	0.139	0.232	0.184	0.167	0.167
	Lead (Pb)-Dissolved (mg/L)	0.000291	0.000440	0.000325	0.000333	0.000321
	Lithium (Li)-Dissolved (mg/L)	0.00356	0.00341	0.00356	0.00370	0.00360
	Magnesium (Mg)-Dissolved (mg/L)	5.24	6.77 ^{DTC}	6.13	6.06	5.99
	Manganese (Mn)-Dissolved (mg/L)	0.0152	0.185 ^{DTC}	0.0623	0.0762	0.0765
	Molybdenum (Mo)-Dissolved (mg/L)	0.000483	0.000481 ^{DTC}	0.000492	0.000484	0.000490
	Nickel (Ni)-Dissolved (mg/L)	<0.00050	0.00501 ^{DTC}	0.00168	0.00189	0.00186
	Phosphorus (P)-Dissolved (mg/L)	<0.30	<0.30	<0.30	<0.30	<0.30
	Potassium (K)-Dissolved (mg/L)	0.567	0.582	0.593	0.593	0.596

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

		Sample ID	L1491715-6	L1491715-7	L1491715-8	L1491715-9	L1491715-10
		Description	Water	Water	Water	Water	Water
		Sampled Date	21-JUL-14	21-JUL-14	21-JUL-14	21-JUL-14	21-JUL-14
		Sampled Time	14:32	14:51	17:18	16:57	17:45
		Client ID	X10	X3A	R10	NF1	FIELD BLANK
Grouping	Analyte						
WATER							
Total Metals	Nickel (Ni)-Total (mg/L)		0.00118	0.00119	<0.00050	0.00075	<0.00050
	Phosphorus (P)-Total (mg/L)		<0.30	<0.30	<0.30	<0.30	<0.30
	Potassium (K)-Total (mg/L)		0.725	0.711	0.578	0.598	<0.050
	Selenium (Se)-Total (mg/L)		0.00026	0.00023	0.00026	0.00024	<0.00010
	Silicon (Si)-Total (mg/L)		4.40	4.36	4.88	4.77	<0.050
	Silver (Ag)-Total (mg/L)		0.000010	<0.000010	<0.000010	0.000011	<0.000010
	Sodium (Na)-Total (mg/L)		1.99	2.01	1.98	1.98	<0.050
	Strontium (Sr)-Total (mg/L)		0.144	0.133	0.113	0.112	<0.00020
	Thallium (Tl)-Total (mg/L)		<0.000010	<0.000010	<0.000010	<0.000010	<0.000010
	Tin (Sn)-Total (mg/L)		<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Titanium (Ti)-Total (mg/L)		<0.010	<0.010	<0.010	<0.010	<0.010
	Uranium (U)-Total (mg/L)		0.00141	0.00139	0.00110	0.00112	<0.000010
	Vanadium (V)-Total (mg/L)		<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Zinc (Zn)-Total (mg/L)		0.127	0.136	0.0052	0.0119	<0.0030
	Zirconium (Zr)-Total (mg/L)		<0.00080	<0.00080	<0.00080	<0.00080	<0.00080
Dissolved Metals	Dissolved Metals Filtration Location		FIELD	FIELD	FIELD	FIELD	FIELD
	Aluminum (Al)-Dissolved (mg/L)		0.0066	0.0072	0.0102	0.0082	<0.0010
	Antimony (Sb)-Dissolved (mg/L)		<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Arsenic (As)-Dissolved (mg/L)		0.00037	0.00043	0.00061	0.00068	<0.00010
	Barium (Ba)-Dissolved (mg/L)		0.0510	0.0503	0.0480	0.0509	<0.000050
	Beryllium (Be)-Dissolved (mg/L)		<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Bismuth (Bi)-Dissolved (mg/L)		<0.00050	<0.00050	<0.00050	<0.00050	<0.00050
	Boron (B)-Dissolved (mg/L)		<0.010	<0.010	<0.010	<0.010	<0.010
	Cadmium (Cd)-Dissolved (mg/L)		0.000067	0.000096	0.000011	0.000020	<0.000010
	Calcium (Ca)-Dissolved (mg/L)		30.9	30.0	26.2	26.5	<0.020
	Chromium (Cr)-Dissolved (mg/L)		<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Cobalt (Co)-Dissolved (mg/L)		0.00034	0.00047	<0.00010	0.00021	<0.00010
	Copper (Cu)-Dissolved (mg/L)		0.00052	0.00051	0.00049	0.00049	<0.00020
	Iron (Fe)-Dissolved (mg/L)		0.293	0.157	0.114	0.249	<0.010
	Lead (Pb)-Dissolved (mg/L)		0.000162	0.000239	0.000095	0.000207	<0.000050
	Lithium (Li)-Dissolved (mg/L)		0.00295	0.00301	0.00331	0.00349	<0.00050
	Magnesium (Mg)-Dissolved (mg/L)		7.10	6.49	5.21	5.53	<0.0050
	Manganese (Mn)-Dissolved (mg/L)		0.0505	0.0556	0.0164	0.0866	<0.000050
	Molybdenum (Mo)-Dissolved (mg/L)		0.000449	0.000401	0.000490	0.000499	<0.000050
	Nickel (Ni)-Dissolved (mg/L)		0.00114	0.00114	<0.00050	0.00059	<0.00050
	Phosphorus (P)-Dissolved (mg/L)		<0.30	<0.30	<0.30	<0.30	<0.30
	Potassium (K)-Dissolved (mg/L)		0.703	0.717	0.539	0.587	<0.050

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID	L1491715-11	Water	TRAVEL BLANK	
Grouping	Analyte				
WATER					
Total Metals	Nickel (Ni)-Total (mg/L)	<0.00050			
	Phosphorus (P)-Total (mg/L)	<0.30			
	Potassium (K)-Total (mg/L)	<0.050			
	Selenium (Se)-Total (mg/L)	<0.00010			
	Silicon (Si)-Total (mg/L)	<0.050			
	Silver (Ag)-Total (mg/L)	<0.000010			
	Sodium (Na)-Total (mg/L)	<0.050			
	Strontium (Sr)-Total (mg/L)	<0.00020			
	Thallium (Tl)-Total (mg/L)	<0.000010			
	Tin (Sn)-Total (mg/L)	<0.00010			
	Titanium (Ti)-Total (mg/L)	<0.010			
	Uranium (U)-Total (mg/L)	<0.000010			
	Vanadium (V)-Total (mg/L)	<0.0010			
	Zinc (Zn)-Total (mg/L)	<0.0030			
	Zirconium (Zr)-Total (mg/L)	<0.00080			
Dissolved Metals	Dissolved Metals Filtration Location				
	Aluminum (Al)-Dissolved (mg/L)				
	Antimony (Sb)-Dissolved (mg/L)				
	Arsenic (As)-Dissolved (mg/L)				
	Barium (Ba)-Dissolved (mg/L)				
	Beryllium (Be)-Dissolved (mg/L)				
	Bismuth (Bi)-Dissolved (mg/L)				
	Boron (B)-Dissolved (mg/L)				
	Cadmium (Cd)-Dissolved (mg/L)				
	Calcium (Ca)-Dissolved (mg/L)				
	Chromium (Cr)-Dissolved (mg/L)				
	Cobalt (Co)-Dissolved (mg/L)				
	Copper (Cu)-Dissolved (mg/L)				
	Iron (Fe)-Dissolved (mg/L)				
	Lead (Pb)-Dissolved (mg/L)				
	Lithium (Li)-Dissolved (mg/L)				
	Magnesium (Mg)-Dissolved (mg/L)				
	Manganese (Mn)-Dissolved (mg/L)				
	Molybdenum (Mo)-Dissolved (mg/L)				
	Nickel (Ni)-Dissolved (mg/L)				
	Phosphorus (P)-Dissolved (mg/L)				
	Potassium (K)-Dissolved (mg/L)				

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID Description Sampled Date Sampled Time Client ID	L1491715-1 Water 21-JUL-14 16:00 NF2B	L1491715-2 Water 21-JUL-14 16:14 NF2	L1491715-3 Water 21-JUL-14 15:45 NF2A	L1491715-4 Water 21-JUL-14 15:21 X2-R	L1491715-5 Water 21-JUL-14 15:11 X2
Grouping	Analyte					
WATER						
Dissolved Metals	Selenium (Se)-Dissolved (mg/L)	0.00025	0.00024	0.00025	0.00023	0.00026
	Silicon (Si)-Dissolved (mg/L)	4.59	5.02	4.82	4.75	4.76
	Silver (Ag)-Dissolved (mg/L)	<0.000010	<0.000010	<0.000010	<0.000010	<0.000010
	Sodium (Na)-Dissolved (mg/L)	1.90	2.00	2.03	1.98	1.99
	Strontium (Sr)-Dissolved (mg/L)	0.108	0.106	0.113	0.110	0.110
	Thallium (Tl)-Dissolved (mg/L)	<0.000010	0.000011	<0.000010	<0.000010	<0.000010
	Tin (Sn)-Dissolved (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Titanium (Ti)-Dissolved (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Uranium (U)-Dissolved (mg/L)	0.000991	0.00102	0.00103	0.00102	0.00102
	Vanadium (V)-Dissolved (mg/L)	<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Zinc (Zn)-Dissolved (mg/L)	0.0065	0.871 ^{DTC}	0.229	0.249	0.250
	Zirconium (Zr)-Dissolved (mg/L)	<0.00080	<0.00080	<0.00080	<0.00080	<0.00080

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

	Sample ID	L1491715-6	L1491715-7	L1491715-8	L1491715-9	L1491715-10
Description	Water	Water	Water	Water	Water	Water
Sampled Date	21-JUL-14	21-JUL-14	21-JUL-14	21-JUL-14	21-JUL-14	21-JUL-14
Sampled Time	14:32	14:51	17:18	16:57	17:45	17:45
Client ID	X10	X3A	R10	NF1	FIELD BLANK	FIELD BLANK
Grouping	Analyte					
WATER						
Dissolved Metals	Selenium (Se)-Dissolved (mg/L)	0.00024	0.00024	0.00025	0.00026	<0.00010
	Silicon (Si)-Dissolved (mg/L)	4.25	4.38	4.77	4.77	<0.050
	Silver (Ag)-Dissolved (mg/L)	<0.000010	<0.000010	<0.000010	<0.000010	<0.000010
	Sodium (Na)-Dissolved (mg/L)	1.95	1.98	1.89	1.98	<0.050
	Strontium (Sr)-Dissolved (mg/L)	0.135	0.135	0.110	0.112	<0.00020
	Thallium (Tl)-Dissolved (mg/L)	<0.000010	<0.000010	<0.000010	<0.000010	<0.000010
	Tin (Sn)-Dissolved (mg/L)	<0.00010	<0.00010	<0.00010	<0.00010	<0.00010
	Titanium (Ti)-Dissolved (mg/L)	<0.010	<0.010	<0.010	<0.010	<0.010
	Uranium (U)-Dissolved (mg/L)	0.00131	0.00134	0.00103	0.00104	<0.000010
	Vanadium (V)-Dissolved (mg/L)	<0.0010	<0.0010	<0.0010	<0.0010	<0.0010
	Zinc (Zn)-Dissolved (mg/L)	0.117	0.131	0.0041	0.0101	<0.0010
	Zirconium (Zr)-Dissolved (mg/L)	<0.00080	<0.00080	<0.00080	<0.00080	<0.00080

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

ALS ENVIRONMENTAL ANALYTICAL REPORT

Grouping	Analyte	Sample ID Description Sampled Date Sampled Time Client ID				
WATER	Dissolved Metals Selenium (Se)-Dissolved (mg/L) Silicon (Si)-Dissolved (mg/L) Silver (Ag)-Dissolved (mg/L) Sodium (Na)-Dissolved (mg/L) Strontium (Sr)-Dissolved (mg/L) Thallium (Tl)-Dissolved (mg/L) Tin (Sn)-Dissolved (mg/L) Titanium (Ti)-Dissolved (mg/L) Uranium (U)-Dissolved (mg/L) Vanadium (V)-Dissolved (mg/L) Zinc (Zn)-Dissolved (mg/L) Zirconium (Zr)-Dissolved (mg/L)	L1491715-11 Water TRAVEL BLANK				

* Please refer to the Reference Information section for an explanation of any qualifiers detected.

Reference Information

QC Samples with Qualifiers & Comments:

QC Type Description	Parameter	Qualifier	Applies to Sample Number(s)
Duplicate	Antimony (Sb)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Beryllium (Be)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Bismuth (Bi)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Copper (Cu)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Iron (Fe)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Lead (Pb)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Phosphorus (P)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Selenium (Se)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Silver (Ag)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Thallium (Tl)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Tin (Sn)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Titanium (Ti)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Duplicate	Vanadium (V)-Dissolved	DLA	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Phosphorus (P)-Total	MS-B	L1491715-1, -10, -11, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Total Organic Carbon	MS-B	L1491715-3
Matrix Spike	Calcium (Ca)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Magnesium (Mg)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Strontium (Sr)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Barium (Ba)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Calcium (Ca)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Magnesium (Mg)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Sodium (Na)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Strontium (Sr)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Calcium (Ca)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Magnesium (Mg)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Strontium (Sr)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Thallium (Tl)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Barium (Ba)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Calcium (Ca)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Magnesium (Mg)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Strontium (Sr)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9
Matrix Spike	Aluminum (Al)-Dissolved	MS-B	L1491715-1, -10, -2, -3, -4, -5, -6, -7, -8, -9

Qualifiers for Individual Parameters Listed:

Qualifier	Description
DLA	Detection Limit adjusted for required dilution
DTC	Dissolved concentration exceeds total. Results were confirmed by re-analysis.
MS-B	Matrix Spike recovery could not be accurately calculated due to high analyte background in sample.
SP	Sample was Preserved at the laboratory

Test Method References:

ALS Test Code	Matrix	Test Description	Method Reference**
ALK-COL-VA	Water	Alkalinity by Colourimetric (Automated)	EPA 310.2
This analysis is carried out using procedures adapted from EPA Method 310.2 "Alkalinity". Total Alkalinity is determined using the methyl orange colourimetric method.			
ANIONS-CL-IC-WR	Water	Chloride by Ion Chromatography	EPA 300.1
This analysis is carried out using procedures adapted from EPA Method 300.1, "Determination of Inorganic Anions by Ion Chromatography", Revision 1.0, April 1999 and from "Determination of Inorganic Anions in Environmental Waters Using a Hydroxide-Selective Column", Application Note 154 v.19, Dionex 2003.			
ANIONS-F-IC-WR	Water	Fluoride by Ion Chromatography	EPA 300.1
This analysis is carried out using procedures adapted from EPA Method 300.1, "Determination of Inorganic Anions by Ion Chromatography", Revision 1.0, April 1999 and from "Determination of Inorganic Anions in Environmental Waters Using a Hydroxide-Selective Column", Application Note 154 v.19, Dionex 2003.			

Reference Information

ANIONS-NO2-IC-WR	Water	Nitrite Nitrogen by Ion Chromatography	EPA 300.1
This analysis is carried out using procedures adapted from EPA Method 300.1, "Determination of Inorganic Anions by Ion Chromatography", Revision 1.0, April 1999 and from "Determination of Inorganic Anions in Environmental Waters Using a Hydroxide-Selective Column", Application Note 154 v.19, Dionex 2003. Nitrate is detected by UV absorbance.			
ANIONS-NO3-IC-WR	Water	Nitrate Nitrogen by Ion Chromatography	EPA 300.1
This analysis is carried out using procedures adapted from EPA Method 300.1, "Determination of Inorganic Anions by Ion Chromatography", Revision 1.0, April 1999 and from "Determination of Inorganic Anions in Environmental Waters Using a Hydroxide-Selective Column", Application Note 154 v.19, Dionex 2003. Nitrate is detected by UV absorbance.			
ANIONS-SO4-IC-WR	Water	Sulphate by Ion Chromatography	EPA 300.1
This analysis is carried out using procedures adapted from EPA Method 300.1, "Determination of Inorganic Anions by Ion Chromatography", Revision 1.0, April 1999 and from "Determination of Inorganic Anions in Environmental Waters Using a Hydroxide-Selective Column", Application Note 154 v.19, Dionex 2003.			
CARBONS-DOC-VA	Water	Dissolved organic carbon by combustion	APHA 5310 TOTAL ORGANIC CARBON (TOC)
This analysis is carried out using procedures adapted from APHA Method 5310 "Total Organic Carbon (TOC)". Dissolved carbon (DOC) fractions are determined by filtering the sample through a 0.45 micron membrane filter prior to analysis.			
CARBONS-TOC-VA	Water	Total organic carbon by combustion	APHA 5310 TOTAL ORGANIC CARBON (TOC)
This analysis is carried out using procedures adapted from APHA Method 5310 "Total Organic Carbon (TOC)".			
EC-MAN-WR	Water	Conductivity by Meter	APHA 2510 (B)
This analysis is carried out using procedures adapted from APHA Method 2510 "Conductivity". Conductivity is determined using an electrode.			
HARDNESS-CALC-VA	Water	Hardness	APHA 2340B
Hardness (also known as Total Hardness) is calculated from the sum of Calcium and Magnesium concentrations, expressed in CaCO ₃ equivalents. Dissolved Calcium and Magnesium concentrations are preferentially used for the hardness calculation.			
IONBALANCE-VA	Water	Ion Balance Calculation	APHA 1030E
Cation Sum, Anion Sum, and Ion Balance (as % difference) are calculated based on guidance from APHA Standard Methods (1030E Checking Correctness of Analysis). Because all aqueous solutions are electrically neutral, the calculated ion balance (% difference of cations minus anions) should be near-zero.			
Cation and Anion Sums are the total meq/L concentration of major cations and anions. Dissolved species are used where available. Minor ions are included where data is present. Ion Balance is calculated as:			
Ion Balance (%) = [Cation Sum-Anion Sum] / [Cation Sum+Anion Sum]			
MET-D-CCMS-VA	Water	Dissolved Metals in Water by CRC ICPMS	APHA 3030 B&E / EPA SW-846 6020A
This analysis is carried out using procedures adapted from "Standard Methods for the Examination of Water and Wastewater" published by the American Public Health Association, and with procedures adapted from "Test Methods for Evaluating Solid Waste" SW-846 published by the United States Environmental Protection Agency (EPA). The procedures may involve preliminary sample treatment by acid digestion, using hotblock, or filtration (APHA 3030B&E). Instrumental analysis is by collision cell inductively coupled plasma - mass spectrometry (modified from EPA Method 6020A).			
MET-T-CCMS-VA	Water	Total Metals in Water by CRC ICPMS	APHA 3030 B&E / EPA SW-846 6020A
This analysis is carried out using procedures adapted from "Standard Methods for the Examination of Water and Wastewater" published by the American Public Health Association, and with procedures adapted from "Test Methods for Evaluating Solid Waste" SW-846 published by the United States Environmental Protection Agency (EPA). The procedures may involve preliminary sample treatment by acid digestion, using hotblock, or filtration (APHA 3030B&E). Instrumental analysis is by collision cell inductively coupled plasma - mass spectrometry (modified from EPA Method 6020A).			
NH3-F-VA	Water	Ammonia in Water by Fluorescence	J. ENVIRON. MONIT., 2005, 7, 37-42, RSC
This analysis is carried out, on sulfuric acid preserved samples, using procedures modified from J. Environ. Monit., 2005, 7, 37 - 42, The Royal Society of Chemistry, "Flow-injection analysis with fluorescence detection for the determination of trace levels of ammonium in seawater", Roslyn J. Weston et al.			
P-T-PRES-COL-VA	Water	Total P in Water by Colour	APHA 4500-P Phosphorus
This analysis is carried out using procedures adapted from APHA Method 4500-P "Phosphorus". Total Phosphorus is determined colourimetrically after persulphate digestion of the sample.			
PH-MAN-WR	Water	pH by Meter	APHA 4500-H (B)
"This analysis is carried out using procedures adapted from APHA Method 4500-H ""pH Value"". The pH is determined in the laboratory using a pH electrode."			
TDS-CALC-VA	Water	TDS (Calculated)	APHA 1030E (20TH EDITION)
This analysis is carried out using procedures adapted from APHA 1030E "Checking Correctness of Analyses".			
TSS-LOW-WR	Water	Total Suspended Solids by Grav. (1 mg/L)	APHA 2540 D

Reference Information

This analysis is carried out using procedures adapted from APHA Method 2540 "Solids". Solids are determined gravimetrically. Total Suspended Solids are determined by filtering a sample through a glass fibre filter and drying the filter at 104 degrees celsius.

ZR-D-MS-VA Water Dissolved Zr in Water by ICPMS EPA SW-846 3005A/6020A

This analysis is carried out using procedures adapted from "Standard Methods for the Examination of Water and Wastewater" published by the American Public Health Association, and with procedures adapted from "Test Methods for Evaluating Solid Waste" SW-846 published by the United States Environmental Protection Agency (EPA). The procedures may involve preliminary sample treatment by acid digestion, using either hotblock or microwave oven, or filtration (EPA Method 3005A). Instrumental analysis is by inductively coupled plasma - mass spectrometry (EPA Method 6020A).

ZR-T-MS-VA Water Total Zr in Water by ICPMS EPA SW-846 3005A/6020A

This analysis is carried out using procedures adapted from "Standard Methods for the Examination of Water and Wastewater" published by the American Public Health Association, and with procedures adapted from "Test Methods for Evaluating Solid Waste" SW-846 published by the United States Environmental Protection Agency (EPA). The procedures may involve preliminary sample treatment by acid digestion, using either hotblock or microwave oven, or filtration (EPA Method 3005A). Instrumental analysis is by inductively coupled plasma - mass spectrometry (EPA Method 6020A).

** ALS test methods may incorporate modifications from specified reference methods to improve performance.

The last two letters of the above test code(s) indicate the laboratory that performed analytical analysis for that test. Refer to the list below:

Laboratory Definition Code	Laboratory Location
WR	ALS ENVIRONMENTAL - WHITEHORSE, YUKON, CANADA
VA	ALS ENVIRONMENTAL - VANCOUVER, BRITISH COLUMBIA, CANADA

Chain of Custody Numbers:

1 2

GLOSSARY OF REPORT TERMS

Surrogate - A compound that is similar in behaviour to target analyte(s), but that does not occur naturally in environmental samples. For applicable tests, surrogates are added to samples prior to analysis as a check on recovery.

mg/kg - milligrams per kilogram based on dry weight of sample.

mg/kg wwt - milligrams per kilogram based on wet weight of sample.

mg/kg lwt - milligrams per kilogram based on lipid-adjusted weight of sample.

mg/L - milligrams per litre.

< - Less than.


D.L. - The reported Detection Limit, also known as the Limit of Reporting (LOR).

N/A - Result not available. Refer to qualifier code and definition for explanation.

Test results reported relate only to the samples as received by the laboratory.

UNLESS OTHERWISE STATED, ALL SAMPLES WERE RECEIVED IN ACCEPTABLE CONDITION.

Analytical results in unsigned test reports with the DRAFT watermark are subject to change, pending final QC review.



L1491715-COFC

Report To		Report Format / Distribution			Select Service Level Below (Rush Turnaround Time (TAT) is not available for all tests)												
Company:	EDI	Select Report Format:	<input checked="" type="checkbox"/> PDF	<input checked="" type="checkbox"/> EXCEL	<input checked="" type="checkbox"/> EDD (DIGITAL)	R	<input checked="" type="checkbox"/> Regular (Standard TAT if received by 3 pm - business days)										
Contact:	Meighan Kearns	Quality Control (QC) Report with Report	<input type="checkbox"/> Yes <input type="checkbox"/> No			P	<input type="checkbox"/> Priority (2-4 bus. days if received by 3pm) 50% surcharge - contact ALS to confirm TAT										
Address:	2195 - 2nd Avenue Whitehorse, YT Y1A 3T8	<input type="checkbox"/> Criteria on Report - provide details below if box checked				E	<input type="checkbox"/> Emergency (1-2 bus. days if received by 3pm) 100% surcharge - contact ALS to confirm TAT										
Phone:	867-393-4882	Select Distribution:	<input type="checkbox"/> EMAIL <input type="checkbox"/> MAIL <input type="checkbox"/> FAX			E2	<input type="checkbox"/> Same day or weekend emergency - contact ALS to confirm TAT and surcharge										
Invoice To		Invoice Distribution			Analysis Request												
Same as Report To <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No		Select Invoice Distribution: <input type="checkbox"/> EMAIL <input type="checkbox"/> MAIL <input type="checkbox"/> FAX			Indicate Filtered (F), Preserved (P) or Filtered and Preserved (F/P) below												
Copy of Invoice with Report <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Email 1 or Fax: sjenner@edynamics.com															
Company: EDI		Email 2: adrienne.turcolte@gov.yk.ca															
Contact: S Jenner																	
Project Information		Oil and Gas Required Fields (client use)															
ALS Quote #:	Q38556	Approver ID:	Cost Center:														
Job #:	14-Y-270	GL Account:	Routing Code:														
PO / AFE:		Activity Code:															
LSD:		Location:															
ALS Lab Work Order # (lab use only)		ALS Contact: Sean Sluggett			Sampler: DH + MM												
ALS Sample # (lab use only)	Sample Identification and/or Coordinates (This description will appear on the report)		Date (dd-mmm-yy)	Time (hh:mm)	Sample Type	ALK-COL-VA, P-T-COL-VA, IONBALANCE-V	ANIONS-ALL-IC-WR, TDS-CALC-VA	EC-MAN-WR, PH-MAN-WR	TSS-LOW-WR	CARBONS-TOC-VA, NH3-F-VA	CARBONS-DOC-VA	MET-T-COMMS-VA, ZR-T-MIS-VA	MET-D-COMMS-VA, ZR-D-MIS-VA	HARDNESS-CALC-VA	Number of Containers		
	R10		21-Jul-14	17:18	Water	R	R	R	R	R	R	R	R	5			
	NF1		21-Jul-14	16:57	↓	↓	↓	↓	↓	↓	↓	↓	↓	↓			
	Field Blank		21-Jul-14	17:45	↓	↓	↓	↓	↓	↓	↓	↓	↓	↓			
	Travel Blank				↓	↓	↓	↓	↓	↓	↓	↓	↓	↓			
Drinking Water (DW) Samples¹ (client use)		Special instructions / Specify Criteria to add on report (client Use)															
Are samples taken from a Regulated DW System? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No		Use CH2M_EQUIS for EDD.															
Are samples for human drinking water use? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No																	
SHIPMENT RELEASE (client use)		INITIAL SHIPMENT RECEPTION (lab use only)			FINAL SHIPMENT RECEPTION (lab use only)												
Released by:	Date:	Time:	Received by:	Date:	Time:	Received by:	Date:	Time:	SAMPLE CONDITION AS RECEIVED (lab use only) Frozen: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No Ice packs: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No Custody seal intact: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No Cooling initiated: <input checked="" type="checkbox"/>								
INITIAL COOLER TEMPERATURES (°C): _____		INITIAL COOLER TEMPERATURES (°C): _____			FINAL COOLER TEMPERATURES (°C): _____												